



NCL Method ITA-8.1

Chemotaxis Assay

Nanotechnology Characterization Laboratory
Frederick National Laboratory for Cancer Research
Leidos Biomedical Research, Inc.
Frederick, MD 21702
(301) 846-6939
ncl@mail.nih.gov
<http://www.ncl.cancer.gov>



This protocol assumes an intermediate level of scientific competency with regard to techniques, instrumentation, and safety procedures. Rudimentary assay details have been omitted for the sake of brevity.

Method written by:

Timothy M. Potter¹,

Edward Cedrone¹,

Barry W. Neun¹

Marina A. Dobrovolskaia^{1,*}

1 - Nanotechnology Characterization Lab., Frederick National Laboratory for Cancer Research,
Frederick, MD 21702

*- address correspondence to: marina@mail.nih.gov

1. Introduction

This document provides a protocol for rapid, quantitative measure of the chemoattractant capacity of a nanoparticulate material. Leukocyte recruitment is a central component of the inflammatory process, both in physiological host defense and in a range of prevalent disorders with an inflammatory component. In response to a complex network of pro-inflammatory signaling molecules (including cytokines, chemokines and prostaglandins), circulating leukocytes migrate from the bloodstream to the site of inflammation.

2. Principles

This assay represents an *in vitro* model, in which promyelocytic leukemia cells HL-60 are separated from control chemoattractant or test-nanoparticles by a 3 μ m filter; the cell migration through the filter is then monitored by detection of cells migrated through the filter into the bottom chamber of the culture plate using the fluorescent dye Calcein AM.

3. Reagents, Materials, Cell Lines, and Equipment

Note: The NCL does not endorse any of the suppliers listed below; these reagents were used in the development of the protocol and their inclusion is for informational purposes only. Equivalent supplies from alternate vendors can be substituted. Please note that suppliers may undergo a name change due to a variety of factors. Brands and part numbers typically remain consistent but may also change over time.

3.1 Reagents

- 3.1.1 Phosphate buffered saline (PBS) (GE Life Sciences, Hyclone, SH30256.01)
- 3.1.2 Bovine serum albumin (BSA) (Sigma-Aldrich, A4503)
- 3.1.3 Fetal bovine serum (FBS) (GE Life Sciences, Hyclone, SH30070.03)
- 3.1.4 RPMI-1640 (GE Life Sciences, HyClone, SH30096.01)
- 3.1.5 Calcein AM (Molecular Probes, C3099)
- 3.1.6 Penicillin streptomycin solution (GE Life Sciences, Hyclone, SV30010)
- 3.1.7 Trypan Blue solution (Gibco, 15250-061)

- 3.1.8 L-glutamine (GE Life Sciences, Hyclone, SH30034.01)
- 3.2 Materials
 - 3.2.1 Pipettes covering the range of 0.05 to 10 mL
 - 3.2.2 Multi-Screen-MIC 96-well filter plates with 3µm membrane (Millipore, MAMIC3S10)
 - 3.2.3 Multi-Screen 96-well culture tray (feeding tray) (Millipore, MAMCS9610)
 - 3.2.4 96 microwell Nunclon optical bottom plates (Nunc, 165305)
 - 3.2.5 Polypropylene tubes, 50 and 15 mL
 - 3.2.6 Multichannel pipettor
- 3.3 Cell Line
 - 3.3.1 HL-60 promyelocytic cells (ATCC, ATCC®-CCL-240™)
- 3.4 Equipment
 - 3.4.1 Centrifuge capable of operating at 400 x g
 - 3.4.2 Refrigerator, 2-8°C
 - 3.4.3 Freezer, -20°C
 - 3.4.4 Cell culture incubator with 5% CO₂ and 95% humidity.
 - 3.4.5 Biohazard safety cabinet approved for level II handling of biological material
 - 3.4.6 Inverted microscope
 - 3.4.7 Vortex
 - 3.4.8 Hemocytometer
 - 3.4.9 Multiwell plate reader capable of operating in fluorescence mode at 485nm and 535nm

4. Reagent and Control Preparation

4.1 Complete RPMI-1640 medium

The complete RPMI medium should contain the following reagents:

20% FBS (heat inactivated)

4 mM L-glutamine

100 U/mL penicillin

100 µg/mL streptomycin sulfate

Store at 2-8°C protected from light for no longer than 1 month. Before use, warm in a water bath.

4.2 Starvation Media (SM)

The starving RPMI medium should contain the following reagents:

0.2 % BSA

4 mM L-glutamine

100U/mL penicillin

100 µg/mL streptomycin sulfate

Store at 2-8°C protected from light for no longer than 1 month. Before use, warm the medium in a water bath.

4.3 Heat-inactivated fetal bovine serum

Thaw a bottle of FBS at room temperature, or overnight at 2-8°C and allow to equilibrate to room temperature. Incubate 30 minutes at 56°C in a water bath mixing every 5 minutes. Single use aliquots may be stored at 2-8°C for up to one month or at a nominal temperature of -20°C indefinitely.

4.4 Positive control

On the day of the experiment, dilute heat-inactivated FBS in serum-free medium supplemented with 0.2% BSA to a final concentration of 20%.

4.5 Negative Control

Use PBS as a negative control. Process this control the same way as your study samples.

4.6 Calcein AM (CAM) working solution

Calcein AM is supplied as a 1 mM solution. Dilute this stock solution in pre-warmed (37°C) 1x PBS to a final concentration of 4 µM (e.g. add 10 µL of stock Calcein AM to 2.49 mL of 1x PBS). The working dilution should be prepared and used *ex tempore*.

4.7 Vehicle Control

Vehicle control is the buffer or media used to formulate test nanomaterials.

Common excipients used in nanoformulations are trehalose, sucrose, and albumin.

However, other reagents and materials are also used alone or in combination.

Vehicle control should match formulation buffer of the test-nanomaterial by both composition and concentration. This control can be skipped if nanoparticles are stored in PBS.

5. Preparation of Study Samples

This assay requires 3 mL of nanoparticles, at 1x the highest final tested concentration dissolved/resuspended in starvation medium. The concentration is selected based on the plasma concentration of the nanoparticle at the intended therapeutic dose. For the purpose of this protocol this concentration is called “theoretical plasma concentration”. Considerations for estimating theoretical plasma concentration were reviewed elsewhere (1) and are summarized in Box 1 below.

Box 1. Example Calculation of Nanoparticle concentration for In Vitro Test

In this example, we assume the mouse dose is known to be 123 mg/kg.

$$\text{human dose} = \frac{\text{mouse dose}}{12.3} = \frac{123 \frac{\text{mg}}{\text{kg}}}{12.3} = 10 \text{ mg/kg}$$

Blood volume constitutes approximately 8% of body weight, (e.g. a 70 kg human has approximately 5.6 L (8% of 70) of blood). This allows us to get a very rough estimation of what the maximum blood concentration may be.

$$\begin{aligned} \text{in vitro concentration}_{\text{human matrix}} &= \frac{\text{human dose}}{\text{human blood volume}} = \frac{70 \text{ kg} \times 10 \frac{\text{mg}}{\text{kg}}}{5.6 \text{ L}} \\ &= \frac{700 \text{ mg}}{5.6 \text{ L}} = 0.125 \text{ mg/mL} \end{aligned}$$

The assay will evaluate 4 concentrations: 10 X (or when feasible 100X, 30X or 5X) of the theoretical plasma concentration, theoretical plasma concentration and two 1:5 serial dilutions of the theoretical plasma concentration. When the intended therapeutic concentration is unknown, the highest final concentration is 1 mg/mL or the highest reasonably achievable concentration. For example, if the final theoretical plasma concentration to be tested is 0.2 mg/mL, then a stock of 2 mg/mL will be prepared and diluted 10-fold (0.2 mg/mL), followed by two 1:5 serial

dilutions (0.04 and 0.008 mg/mL). Use 150 μ L of each of these samples per well. Each nanoparticle concentration is plated 3 times.

6. HL-60 Cell Preparation

HL-60 is a promyelocytic cell line derived by S.J. Collins, et al. from a patient with acute promyelocytic leukemia (2). Cultures can be maintained by the addition of fresh medium or replacement of medium. Alternatively, cultures can be established by centrifugation with subsequent resuspension at 1×10^5 viable cells/mL. Do not allow cell concentration to exceed 1×10^6 cells/mL. Maintain cell density between 1×10^5 and 1×10^6 viable cells/mL.

- 6.1 Expand cells until they are approximately 80-90% confluent (~3-5 days before experiment). Two days before the experiment, feed the cells following regular maintenance procedures.
- 6.2 One day before the experiment count cells using trypan blue. If the cell viability is at least 90%, pellet cells for 5 minutes at 400 x g in a 15 mL tube.
- 6.3 Resuspend cells in Starvation Medium and incubate overnight (16 - 18h) at 37°C in a humidified incubator (95% air, 5% CO₂).
- 6.4 On the day of experiment, count cells using trypan blue and adjust concentration to 1×10^6 **viable** cells/mL in Starvation Medium.

7. Experimental Procedure

The procedure described below is based on reference 3.

- 7.1 Insert the filter plate into the feeding tray and set it aside.
- 7.2 Add 150 μ L of starvation medium, positive control, negative control and test-nanomaterial in starvation medium into fresh feeding tray. Avoid generating bubbles while adding solutions to wells. Plate additional test-nanomaterial wells for use as cell-free controls (refer to plate map in Appendix).
- 7.3 Add 50 μ L of cell suspension (prepared in step 6.4) per well of the Multi-Screen filter plate (50,000 cells per well) from Step 7.1. Avoid generating bubbles while adding cells to wells. Add 50 μ L of media to the cell-free control wells.

- 7.4 Gently assemble Multi-Screen filter plate (step 7.3) and feeding tray containing controls and test particles (step 7.2). This is the **Assay Plate**. An example of the assay plate template is shown in the Appendix. Avoid shaking or tilting plates as it will disturb concentration gradient.
- 7.5 Cover the plate and incubate for 4 h at 37°C in a humidified incubator (5% CO₂, 95% air). During incubation, pre-warm PBS to 37°C and equilibrate the calcein AM to room temperature.
- 7.6 Prepare a working solution of Calcein AM (CAM) as described in section 4.7 of the method.
- 7.7 After the 4 h incubation, remove the chemotaxis assay plate from the incubator. As before, avoid shaking or tilting the plate.
- 7.8 Gently remove the Multi-Screen filter plate and discard it.
- 7.9 Add 50 µL of 1x PBS and 50 µL of CAM working solution to appropriate wells, and 150 µL of 1x PBS plus 50 µL of CAM working solution to the reagent background control wells on the feeding tray as outlined in the example template shown in the Appendix. This is the **Calcein Plate**. Incubate the plate for 1 h at 37°C.
- 7.10 Transfer 180 µL of solution from the Calcein Plate to corresponding wells on a Nunc optical bottom plate and read the Nunc plate on a fluorescent plate reader at 485 nm excitation/535 nm emission.

8. Calculations

- 8.1 A percent coefficient of variation (CV) should be calculated for each control or test according to the following formula: $\%CV = SD/Mean \times 100\%$
- 8.2 Background Chemotaxis =
$$\text{Mean } FU_{SM/CAM \text{ wells}} - \text{Mean } FU_{SM/PBS \text{ wells}} - \text{Mean } FU_{\text{reagent background control wells}}$$
- 8.3 Sample Chemotaxis =
$$\text{Mean } FU_{TS/CAM \text{ wells}} - \text{Mean } FU_{SM/PBS} - \text{Mean } FU_{\text{reagent background control wells}}$$
- 8.4 Comparison of sample chemotaxis to background chemotaxis is performed to evaluate chemotactic potential of test material. In general, fold chemotaxis induction equal to or greater than 2 is considered positive. Statistical methods

such as Student's t-test should be applied to evaluate the significance of difference between background chemotaxis and that of test-sample.

9. Acceptance Criteria

- 9.1 %CV for each control and test sample should less than 30%.
- 9.2 If two of three replicates of positive control or negative control fail to meet acceptance criterion described in 9.1, the assay should be repeated.
- 9.3 Within the acceptable assay, if two of three replicates of unknown sample fail to meet acceptance criterion described in 9.1, this unknown sample should be re-analyzed.

10. References

- 1. Dobrovolskaia MA, McNeil SE. Understanding the correlation between in vitro and in vivo immunotoxicity tests for nanomedicines. J Control Release. 2013 Dec 10;172(2):456-66
- 2. Gallagher R, et al. Characterization of the continuous, differentiating myeloid cell line (HL-60) from a patient with acute promyelocytic leukemia. Blood 54: 713-733, 1979.
- 3. Kamath L. Multi-Screen-MIC Application note. Lit# AN1060EN00, Millipore Life Science Division.

11. Abbreviations

BSA	bovine serum albumin
CAM	Calcein AM
CV	coefficient of variation
FBS	fetal bovine serum
FU	fluorescence units
PBS	phosphate buffered saline
RBC	Reagent background control
RPMI	Roswell Park Memorial Institute
SD	standard deviation

SM starvation media
VC vehicle control

12. Appendix

Example Assay Plate Map

	1	2	3	4	5	6	7	8	9	10	11	12
A	SM	SM	SM	SM	SM	SM	NC	NC	NC	PC	PC	PC
B	SM	SM	SM	SM	SM	SM	NC	NC	NC	PC	PC	PC
C	SM	SM	SM	SM	SM	SM	NC	NC	NC	PC	PC	PC
D	TS 0.008 mg/mL	TS 0.008 mg/mL	TS 0.008 mg/mL	TS 0.04 mg/mL	TS 0.04 mg/mL	TS 0.04 mg/mL	TS 0.2 mg/mL	TS 0.2 mg/mL	TS 0.2 mg/mL	TS 1.0 mg/mL	TS 1.0 mg/mL	TS 1.0 mg/mL
E	TS 0.008 mg/mL	TS 0.008 mg/mL	TS 0.008 mg/mL	TS 0.04 mg/mL	TS 0.04 mg/mL	TS 0.04 mg/mL	TS 0.2 mg/mL	TS 0.2 mg/mL	TS 0.2 mg/mL	TS 1.0 mg/mL	TS 1.0 mg/mL	TS 1.0 mg/mL
F	TS 0.008 mg/mL	TS 0.008 mg/mL	TS 0.008 mg/mL	TS 0.04 mg/mL	TS 0.04 mg/mL	TS 0.04 mg/mL	TS 0.2 mg/mL	TS 0.2 mg/mL	TS 0.2 mg/mL	TS 1.0 mg/mL	TS 1.0 mg/mL	TS 1.0 mg/mL
G	VC	VC	VC									
H	NO CELLS TS 0.008 mg/mL	NO CELLS TS 0.008 mg/mL	NO CELLS TS 0.008 mg/mL	NO CELLS TS 0.04 mg/mL	NO CELLS TS 0.04 mg/mL	NO CELLS TS 0.04 mg/mL	NO CELLS TS 0.2 mg/mL	NO CELLS TS 0.2 mg/mL	NO CELLS TS 0.2 mg/mL	NO CELLS TS 1.0 mg/mL	NO CELLS TS 1.0 mg/mL	NO CELLS TS 1.0 mg/mL

This is the template for the feeding tray. Wells in rows A-F in the upper filter plate receive 50,000 cells in 50 μ L. Row H are the cell-free test samples; they do not receive cells.

SM: Starvation Medium; NC: Negative Control; PC: Positive Control; TS: Test Sample; VC – vehicle control

Example Calcein Plate Map

	1	2	3	4	5	6	7	8	9	10	11	12
A	PBS	PBS	PBS	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM
B	PBS	PBS	PBS	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM
C	PBS	PBS	PBS	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM
D	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM
E	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM
F	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM
G	CAM	CAM	CAM	RBC	RBC	RBC	RBC	RBC	RBC	RBC	RBC	RBC
H	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM	CAM

PBS: 50 μ L PBS;

CAM: 50 μ L Calcein AM;

RBC: Reagent Background Control, 150 μ L PBS + 50 μ L Calcein AM